Abstract P132 Figure 1 RFcsa taken with a linear probe (image on left) and curvilinear probe (image on right) in the same individual at 2/3.

Conclusion These data demonstrate that both linear and curvilinear probes can be used to acquire accurate RFcsa measurements. Furthermore, “splicing” the images from the LUP, when a CUP is not available is a justified method to assess RFcsa. This method should be considered for RFcsa image acquisition in obese patients.

P133 LUNG CLEARANCE INDEX IS A REPRODUCIBLE AND SENSITIVE MEASURE OF AIRWAYS DISEASE IN BRONCHIECTASIS

doi:10.1136/thoraxjnl-2011-201054c.133

¹S Rowan, ¹J Bradley, ¹M Ennis, ²A Horsley, ³N Bell, ⁴P Gustafsson, ¹J S Elborn. ¹Queens University Belfast, Belfast, N Ireland; ²University of Manchester, Manchester, UK; ³Bristol Royal Infirmary, Bristol, UK; ⁴Central Hospital, Skovde, Sweden

Introduction Lung clearance index (LCI) is a measure of ventilation in homogeneity derived from multiple breath washout (MBW). Although FEV₁ is commonly used to assess severity of airway disease and response to therapy, it is insensitive to small airways disease and is often within normal range in bronchiectasis (BE) not caused by Cystic Fibrosis (CF) until disease is well established. In CF, LCI is more sensitive than FEV₁ in detecting airways abnormalities and is currently used as an outcome measure in clinical trials. In BE, there is a need to find a sensitive outcome measure that is responsive to interventions, particularly in those with mild disease.

Objective To assess within and between visit repeatability of LCI and determine the relationship between FEV₁ and LCI in stable BE.

Methods Inclusion criteria: HRCT diagnosis of BE within the last 5 years; clinically stable (no infective symptoms for >4 weeks); no genetic or clinical features of CF. Participants attended for two visits, 2 weeks apart. At each visit they performed MBW in triplicate, using 0.2% sulphur hexafluoride and a modified Innocor device. LCI was derived from the mean of at least 2 acceptable washouts. Spirometry was performed to ATS/ERS standards.

Results 14 patients (8M/6F) attended for two visits. The mean (SD) age was 60.5 (15.4) yrs. Mean (SD) FEV₁ % predicted was 87.1 (18.6), range (44–117). Mean (SD) LCI was 9.4 (2.0) on visit 1 and 9.4 (1.9) on visit 2 (normal <7.5). The intra-visit coefficient of variation (CV) was 4.7 % (3 measures). Between visit repeatability of LCI was 0.54 (SD of variance between visits). LCI negatively correlated with FEV₁ ($r = -0.69$, $p < 0.001$). Sensitivity of LCI and FEV₁ for the diagnosis of bronchiectasis by CT was 71% and 29% respectively.

Conclusions This is the first report of LCI in non-CF BE. LCI is a more sensitive test of lung function than FEV₁ and is abnormal in the majority of people with BE who have a normal FEV₁. LCI has good intra-visit and between visit repeatability. Across a range of FEV₁ there is a strong relationship between LCI and FEV₁.

P134 VALIDATING STRUCTURED LIGHT PLETHYSMOGRAPHY (SLP) AS A NON-INVASIVE METHOD OF MEASURING LUNG FUNCTION WHEN COMPARED TO SPIROMETRY

doi:10.1136/thoraxjnl-2011-201054c.134

¹S Alimohamed, ¹K Prosser, ¹C Weerasuriya, ²R Iles, ³J Cameron, ³J Lasenby, ⁴C Fogarty. ¹School of Clinical Medicine, University of Cambridge, Addenbrookes Hospital, Cambridge, UK; ²Department of Respiratory Paediatrics, Cambridge University Hospitals, Addenbrookes Hospital, Cambridge, UK; ³Signal Processing and Communications Laboratory, Department of Engineering, University of Cambridge, Cambridge, UK; ⁴Department of Statistics, Harvard University, Cambridge, Massachusetts, USA

Background Structured Light Plethysmography (SLP) is a recently developed technology for non-invasive and entirely non-contact monitoring of lung (respiratory) function. The system projects a structured light grid onto the thoraco-abdominal surface of the subject, which is imaged by two cameras giving a dynamic 3D reconstruction of the surface as the subject breathes. From this data we can infer changes in chest/abdomen volume over time, allowing us to extract parameters and generate curves (eg, Volume-Time, Flow-Time, Flow-Volume curves) directly comparable to conventional spirometry. SLP therefore hopes to provide an inexpensive replacement for conventional spirometry, which is an invasive methodology unusable in a number of patient classes (eg, neonates, young children, intensive care patients etc). This study tests the validity of SLP in terms of reproducibility, repeatability and position dependence, as compared to conventional spirometry (Pneumatach); by comparison of ventilation parameters extracted from both technologies.

Methods SLP and Pneumatach spirometry were used simultaneously to capture 120 datasets from 10 randomly chosen adult subjects. Each complete dataset contained tidal breathing and forced expiratory manoeuvres, in both sitting and standing positions. Operator-dependence (reproducibility) was tested by collecting data sets from each subject using three different operators. Repeatability was tested by collecting the data from each subject once, and then again after a 40 min break. Tidal Inspiratory Time (TI) parameters were extracted from the results and the data analysed using the paired Student t test.

Results There was no significant difference between TI values obtained from SLP compared to conventional spirometry throughout the study ($n=120$; $p=0.8556$). SLP comparisons of pooled mean TI before, and after a 40 min break were not significant (1.5589 vs 1.5595; $p=0.9938$); similarly, readings in different positions (sitting or standing) were not significantly different. SLP comparison of all three operator pairs (1vs2, 1vs3 and 2vs3), were not significant ($p=0.7361$, $p=0.9765$, $p=0.7343$, respectively).

Conclusions SLP measurements are not operator, time or position dependant. Therefore SLP shows a high degree of reproducibility and repeatability; and represents a promising, viable and non-invasive alternative to conventional spirometry.

P135 MAXIMUM INSPIRATORY FLOW MEASURED DIRECTLY WITH AN INSPIRATORY FLOW METRE COMPARED WITH MEASUREMENTS FROM FLOW VOLUME LOOP TRACES

doi:10.1136/thoraxjnl-2011-201054c.135

M C P Apps, M Skipper, J Skipper, S Basford, V Bryant, J Saines, T Andrews. Queen's Hospital, Romford, UK

Introduction Low Maximum Inspiratory Flow is characteristic of patients with muscle weakness, Laryngeal dysfunction, or Extra-thoracic airway obstruction. We have been exploring the use of an Inspiratory Flow Metre (In Check Dial- Clement Clarke International) to help in the identification of those patients who need more detailed investigation.