REFERENCES

1. Ahmad N, et al. Acute hypercapnic respiratory failure (AHRF): Looking at long-term mortality, prescription of long-term oxygen therapy and chronic non-invasive ventilation (NIV). *Clin Med (Lond)* 2012;12(2):188.
2. Kaul S, et al. Non-invasive ventilation (NIV) in the clinical management of acute COPD in 233 UK hospitals: Results from the RCP/BTS 2003 national COPD audit. *COPD* 2009;6(3):171–6.

Characterisation of lung disease with imaging and physiology

P134 AN EVALUATION OF A NEW LUNG FUNCTION TEST: TLNO IN HEALTHY SUBJECTS

A Ijaz, J Hull. Imperial College London, London, UK

10.1136/thoraxjnl-2017-210983.276

Background and Aims TLNO, transfer factor for nitric oxide, is a pulmonary function test of gas transfer. The test, in combination with carbon monoxide, allows for calculation of the alveolar membrane diffusing capacity and red blood cell conductance. This allows physicians to recognise physiologically where issues with gas transfer arise. Currently, there are very few papers looking at TLNO reproducibility in healthy subjects and none aiming with a target of 5% repeatability between efforts. This study attempts to look at the reproducibility of TLNO over 10 sessions (7 weeks total) with an intra-session repeatability of 5%. In addition, comparison of TLCO measurements between 10 and 5 s breath holds are made.

Methods 14 normal subjects were recruited and a baseline spirometry was taken and height, weight, age and sex were recorded. Subjects were asked to perform a TLCO test with 10 s breath hold followed by 10 repeated sessions of the TLNO test on different days. Measurements within 5% of each other were considered acceptable repeated Results in one session. A Bland-Altman plot and regression line were constructed to compare TLCO measures between different breath hold times. One-way repeated measures ANOVA, measurement error values, intra and inter-session variability were calculated for TLNO and TLCO recordings obtained over the 10 repeated sessions.

Results Bland-Altman plot revealed no statistically significant ($p=0.783$, $p\geq 0.05$) difference between TLCO breath hold times. Coefficient of determination from the regression line, $r^2=0.860$. Repeated measures ANOVA revealed no significant difference for TLNO and TLCO measurements over time at $p=0.374$ and $p=0.842$ ($p\geq 0.05$) respectively. Intra-session and inter-session variability for TLNO were calculated as 15.02 ml/min/mmHg and 16.12 ml/min/mmHg respectively. TLCO intra-session and inter-session variability were 4.30 ml/min/mmHg and 3.70 ml/min/mmHg. We have shown that TLCO values recorded with the shorter 5 s breath hold agree with the conventional 10 s technique. Over a 7 week period TLNO and TLCO do not change significantly and calculated session variability is consistent with ERS/ATS guidelines.

Conclusion TLNO is a highly reproducible test over a 7 week period and a shortened breath hold in healthy people provides the same values as the traditional 10 s breath hold for TLCO.

P135

THE PREVALENCE OF UNDIAGNOSED COPD IN PATIENTS WITH AN ABDOMINAL AORTIC ANEURYSM AND ITS IMPACT ON CARDIOPULMONARY EXERCISE TESTS

L Archer, E Parkes, J Shakespeare, DP Parr. University Hospitals Coventry and Warwickshire, Coventry, UK

10.1136/thoraxjnl-2017-210983.277

Introduction Cigarette smoking is a known contributing factor to the development of chronic obstructive pulmonary disease (COPD), and to the formation of an aneurysm. In addition COPD is the most important risk factor associated with abdominal aortic aneurysms (AAA). Studies¹ have observed the relationship of COPD and AAA, demonstrating a prevalence in COPD patients of 3.7% and a 1.22–1.78 fold increased risk of AAA when compared to those without COPD.²

Aims Our aim was to identify the prevalence of undiagnosed airflow obstruction in patients attending for CPET for preoperative assessment for AAA.

Methods Data from 122 patients (108 male), median age 75 years (range 65–90) with an AAA of 5 cm or more, attending between September 2014 and May 2016 were included. Spirometry, CPET, smoking history, BMI and current medication were all analysed. The patient's clinical records were reviewed to establish any previous respiratory diagnosis.

Results 17/122 (14%) patients had a known diagnosis of COPD however 32/122 (26%) had airflow obstruction on spirometry, with only 12 of these having a diagnosis of COPD. 5 patients with a diagnosis of COPD did not demonstrate airflow obstruction. There were significant differences between those with airflow obstruction and those without for breathing reserve 20.8 versus 37.95% ($p=0.0002$), anaerobic threshold 11.0 versus 12.30 ml/min/kg ($p=0.0073$), Peak VO2 15.1 versus 16.8 ml/min/kg ($p=0.0018$), smoking history 49 versus 30 pack years ($p=0.0069$) and BMI 26.0 versus 28.1 kg/m² ($p=0.0024$); respectively.

Conclusion The Results confirmed that a significant proportion of our patients had previously undiagnosed/unrecognised airflow obstruction. As would be expected, patients with airflow obstruction had an increased pack year smoking history and a decreased BMI when compared to those without. Airflow obstruction also resulted in a decreased peak VO2 and a worsening of ventilatory limitation which has the potential to influence surgical decision making.

REFERENCES

1. Aggarwal S, Qamar A, Sharma V, Sharma A. Abdominal aortic aneurysm: a comprehensive review. *Exp Clin Cardiol* 2011;16(1):11–15.
2. Flessenkaemper IH, Lodenkemper R, Roll S, Enke-Melzer K, Wurps H, Bauer TT. Screening of COPD patients for abdominal aortic aneurysm. *Int J Chron Obstruct Pulmon Dis* 2015, Jun 10;10:1085–91.

P136

NON-INVASIVE ASSESSMENT OF DIAPHRAGM CONTRACTILITY USING SURFACE MECHANOMYOGRAPHY IN HEALTHY SUBJECTS

¹M Lozano-García, ²L Sarlabous, ³J Moxham, ³GF Rafferty, ²R Jané, ⁴CJ Jolley. ¹Centro de Investigación Biomédica en Red en Bioingeniería, Biomateriales y Nanomedicina, Barcelona, Spain; ²Institute for Bioengineering of Catalonia (IBEC), The Barcelona Institute of Science and Technology, Barcelona, Spain; ³Division of Asthma Allergy and Lung Biology, King's College London, King's Health Partners, London, UK; ⁴Centre of Human and Aerospace Physiological Sciences, King's College London, King's Health Partners, London, UK

10.1136/thoraxjnl-2017-210983.278

Introduction The current gold standard assessment of diaphragm contractility involves invasive measures of transdiaphragmatic pressure (Pdi) (ATS/ERS 2002). Surface mechanomyography (sMMG) is a non-invasive measure of muscle fibre vibration during contraction. Sarlabous et al (ERJ 2015) reported a high correlation between diaphragm sMMG (sMMGdi) amplitude and inspiratory mouth pressure, a measure of global respiratory muscle function. To further validate the technique, the relationship between sMMGdi and Pdi was examined in this study, hypothesising that there would be a close relationship between sMMGdi and Pdi in healthy subjects.

Methods Pdi and sMMGdi (right lateral chest) were measured in 12 healthy subjects (6 male, age 33 (30–38) years, BMI 22.2 (20.6–24.2) kg/m²) during an increasing inspiratory threshold loading protocol. sMMGdi signals were analysed using fixed sample entropy (fSampEn). Mean and peak values, and the area under the curve of inspiratory Pdi and fSampEn sMMGdi, were calculated and compared.

Abstract P136 Figure 1 Relationship between the area under the curve (A), the mean value (B), and the peak value, and (C) of Pdi and fSampEn sMMGdi, during an increasing inspiratory threshold loading protocol. In each subfigure, the mean and standard deviation values of all healthy subjects are shown at rest and at 12%, 24%, 36%, 48%, and 60% of maximum mouth inspiratory pressure. In subfigure A, the entropy-time product (ETP) and the pressure-time product (PTP) represent the area under the fSampEn sMMGdi and Pdi curves, respectively, multiplied by the respiratory rate.

Results Strong correlations between the non-invasive sMMGdi and the invasive Pdi measures were observed for all three parameters (figure 1).

Conclusion The strong correlation between measures of sMMGdi and Pdi in healthy subjects suggests that sMMGdi could provide a clinically applicable noninvasive index of respiratory muscle contractility in patients with respiratory muscle weakness for diagnosis and monitoring.

P137 19 F-MRI OF INHALED PERFLUOROPROPANE GAS: A NOVEL APPROACH TO VENTILATION IMAGING

¹B Pippard, ¹M Neal, ²P Dutta, ²AJ Simpson, ¹P Thelwall. ¹Newcastle Magnetic Resonance Centre, Newcastle University, Newcastle upon Tyne, UK; ²Institute of Cellular Medicine, Newcastle University, Newcastle upon Tyne, UK

10.1136/thoraxjnl-2017-210983.279

Introduction Current measures of pulmonary ventilation are beset by limitations. For example, spirometry provides no regional or anatomical information regarding lung function, while both CT and V/Q scans incur ionising radiation doses, restricting serial use. Hyperpolarised gas MRI (e.g., using ³He and ¹²⁹Xe) enables direct assessment of airway structure and function without recourse to ionising radiation, but is confined to centres with access to specialised hyperpolarising equipment and expertise. A novel approach, involving ¹⁹F-MRI of inhaled perfluoropropane gas, has recently been described in humans,^{1,2} offering an alternative to hyperpolarisation with scope for translation to clinical practice.

Aim We assessed the feasibility of using inhaled perfluoropropane gas to image pulmonary ventilation properties in a group of healthy volunteers.

Methods 17 participants (10 M, 7 F; aged 21–52) provided written informed consent, and were screened for normal lung function using standard spirometry. Participants were invited to attend two MRI scanning sessions, during which they inhaled a 79% perfluoropropane/21% oxygen gas mixture on up to three occasions. Gas inhalations lasted <1 min, typically involving 3–5 deep breaths followed by breath-hold. MRI scans of inhaled perfluoropropane were acquired using a Philips Achieva 3T scanner and a designated receiver coil tuned to ¹⁹F frequency. Heart rate and oxygen saturations were monitored throughout.

Results In total, 94 gas inhalations were performed across 17 participants. Ventilation images were obtained within a single breath-hold, demonstrating homogeneous gas distribution throughout the lungs (figure 1). Inhalation of the gas mixture was well tolerated with no significant adverse events, other than a transient (seconds) fall in SpO₂ (89%) following one breath-hold. This resolved spontaneously, and was not considered clinically significant.

Conclusions ¹⁹F-MRI of inhaled perfluoropropane gas represents a novel approach to ventilation imaging, with potential for high quality image acquisition within a single breath-hold. Crucially, this technique can be implemented on MRI scanners with considerably less additional hardware than required for hyperpolarised gas MRI. Ongoing optimisation of scan protocols will enable further improvements in both spatial and temporal resolution, providing a platform for future clinical application.